



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Eurofins Central Laboratory*, using *Bio-Rad Variant II* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: Variant II	Calibrator Lot: 64232925, 64232927	Column Lot: NU70415F
Reagent Lot: 64284285, 64284286, 64279163	Calibrator Assigned Values: 5.6%, 10.1%	

Date of Certification: June 1, 2020

Certification Expires: June 1, 2021

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor