



Certificate of Traceability

Level I Laboratory Certification

This certifies that ***Eurofins Central Laboratory China Co. Ltd***, using ***Bio-Rad Variant II*** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **Reference Material Institute for Clinical Chemistry Standards (ReCCS) ASRL#1**

The system evaluated was:

Instrument: Variant II	Calibrator Lot: 64232889, 64232907	Column Lot: CN70415
Reagent Lot: AA80445, AA80446, 64259209	Calibrator Assigned Values: 5.5%, 10.2%	

Date of Certification: June 1, 2020

Certification Expires: June 1, 2021

Daniel Sacks

NGSP Steering Committee Chair

Randie R Little PhD.

NGSP Network Coordinator

Violeta G. Raneva, Ph.D.

SRL director/ supervisor